

Further Investigation of the Performance of the XSD MINICAMS®

Update on a Topic Presented at CWD2014



*Dr. Gary D. Sides, Ms. Claire E. Longo, Ms. Carrie A. Kauffman,
and Mr. Walter L. Waybright*

CWD 2015

June 5, 2015



PCAPP

Pueblo Chemical
Agent-Destruction Pilot Plant

www.peoacwa.army.mil

A PARTNERSHIP FOR SAFE CHEMICAL WEAPONS DESTRUCTION



Acronyms



ABCDF	Aberdeen Chemical Agent Disposal Facility
ACWA	Assembled Chemical Weapons Alternatives
BGCAPP	Blue Grass Chemical-Agent Disposal Facility
DELCD	Dry electrolytic conductivity detector
FAL	Found action level
GC	Gas chromatographic; gas chromatography
H	Chromatographic peak height
H _{CAL}	Chromatographic calibration peak height
HD	Agent mustard
LMQAP	Laboratory and Monitoring Quality Assurance Plan
LOQ	Limit of quantification
n	Exponent used within the MINICAMS; n = 1.00 for a linear detector
nA	Nanoamperes
OICO	O.I. Corporation
P&A	Precision and accuracy
PCAPP	Pueblo Chemical Agent-Destruction Pilot Plant
PM	Preventive maintenance
TAL	Target action level
UIFM	Uncertainty in found mass
VSL	Vapor screening level
XSD	Halogen Specific Detector



MINICAMS[®] with a Halogen Selective Detector (XSD)



- Sensitive with a typical detection limit of < 0.5 ng for agent mustard (HD)
- Selective by a factor of about 8,000 for chlorinated compounds vs. hydrocarbons
- Used for monitoring HD, Lewisite, and other chlorinated agents since about 1996
- Used during the destruction of agent mustard at the Aberdeen Chemical Agent Disposal Facility (ABCDF), 2003-2006
- Will be used during the destruction of agent mustard at the Pueblo Chemical Agent-Destruction Pilot Plant (PCAPP) and at the Blue Grass Chemical Agent-Destruction Pilot Plant (BGCAPP)
- Previously assumed to be a linear detector in demilitarization applications



XSD MINICAMS



ABCDF




MINICAMS[®] is a registered trademark of CMS Research Corporation, Pelham, Alabama.

Requirements for Certification and Validation of an Air Monitoring System and Method*



- **Certification** at U.S. agent destruction sites requires
 - passing a precision-and-accuracy (P&A) study consisting of a series of challenges conducted twice per day over a 4-day period
 - passing an initial baseline study consisting of at least one challenge per day over a 28-day period
- **Validation** requires
 - passing a continuing baseline study consisting of at least one challenge per day evaluated each 28-day period

Study Name: P&A Study



FAL: 0.8945
TAL: 0.7905
LOQ: 0.0288
DETL: 0.0130
DECL: 0.0291
UIFM: 11.38 %
UIAS: 10.48 %

Slope: 0.9932
Y-intercept: 0.0162
Percent Recovery: 100.93 %
Data Points: 48

*ACWA Programmatic Laboratory and Monitoring Quality Assurance Plan



Requirements at the Pueblo Chemical Agent-Destruction Pilot Plant (PCAPP)



- **Certification by**
 - passing a precision-and-accuracy (P&A) study in the range of **0.2 to 2 VSL** for HD (vs. the normal range of 0.5 to 2 VSL)
 - passing statistical requirements for an initial baseline study consisting of at least one challenge per day over a 28-day period
- **Demonstrating a limit of quantification (LOQ) of < 0.2 VSL**
(Note that some decisions at the site will be made on the basis of MINICAMS concentration reports that are >0.2 VSL—versus greater than a true concentration of 0.2 VSL.)
- **Validation by**
 - passing a continuing baseline study consisting of at least one challenge per day evaluated statistically each 28-day period

1.0 VSL = 0.003 mg/m³

Equivalent to 4.05 ng at PCAPP (for a sample flow rate of 450 mL/min and a sample period of 3 min)



Equations Relating Peak Height to Concentration in the MINICAMS Code



The equation used in the MINICAMS software to relate peak height (H) to concentration for calibration at a single challenge concentration (1.00 VSL) is:

$$H = H_{\text{CAL}} (\text{VSL})^n$$

where

H_{CAL} is the average peak height obtained for 1.0-VSL challenges (i.e., for a single-point calibration)

and

n is a coefficient, which is set to 1.00 for an assumed linear response.

For a given peak height, the concentration reported by the MINICAMS is given by:

$$\text{VSL} = (H/H_{\text{CAL}})^{1/n}$$

CERTIFY Results for PCAPP P&A Study

Oct 9-10, 2013, assuming linearity (i.e., $n = 1.00$)




Preliminary study conducted over two days to get an indication of MINICAMS performance

- Calibrated two XSD MINICAMS units using two 1.0-VSL-equivalent injections for each
- A series of 6 challenges of each MINICAMS was conducted twice per day over a 2-day period (0.0, 0.2, 0.75, 1.0, 1.5 and 2.0 VSL) —48 challenges
- **P&A study passed requirements in the PCAPP LQCP, but UIFM (23%) marginal** —raising PCAPP concerns about the robustness of the method
- **PCAPP goal of LOQ < 0.2 VSL was not met**
- **Accuracy less than optimum** —found concentrations high at less than 1.0 VSL and low at greater than 1.0 VSL

Found (reported) versus target concentration (challenge)

Study Name: 9-10 Oct 9366 (210) 9370 (1701)



FAL: 0.6972

TAL: 0.5539

LOQ: 0.3018

DETL: 0.1344

DECL: 0.1929

UIFM: 22.67 %

UIAS: 22.28 %

Slope: 0.8186

Y-intercept: 0.0829


Percent Recovery: 90.15 %

Data Points: 48

$n = 1.00$



Analysis of Initial Response Curves Obtained at PCAPP Yielded an Exponent Value of $n = 0.76$



CERTIFY Results for a PCAPP P&A Study


Nov 20-23, 2013, assuming linearity (i.e., $n = 1.00$)



Conducted in accordance with all requirements in the ACWA LMQAP

- Calibrated two XSD MINICAMS units using two 1.0-VSL injections for each
- A series of 6 challenges of each MINICAMS was conducted twice per day over a 2-day period (0.0, 0.2, 0.75, 1.0, 1.5 and 2.0 VSL) —48 challenges (plus 0.5-VSL challenges)
- 6 different standard solutions; 2 μL injected for each challenge; blind, randomized challenges
- P&A study failed UIFM (27%)—raising PCAPP concerns about the robustness of the method
- PCAPP goal of LOQ < 0.2 VSL was met
- Accuracy less than optimum—found concentrations high at less than 1.0 VSL and low at greater than 1.0 VSL

Study Name: Nov 20-23 without 0.5 with $n=1.0$



FAL: 0.6711

TAL: 0.5412

LOQ: 0.1426

DETL: 0.0559

DECL: 0.0887

UIFM: 26.98 %

UIAS: 22.93 %

Slope: 0.8796

Y-intercept: 0.0395

Percent Recovery: 91.91 %

Data Points: 48

$n = 1.00$



CERTIFY Results for PCAPP P&A Study


Nov 20-23, 2013, assuming non-linearity ($n = 0.76$)



Conducted in accordance with all requirements in the ACWA LMQAP

- Calibrated two XSD MINICAMS units using two 1.0-VSL injections for each
- A series of 6 challenges of each MINICAMS was conducted twice per day over a 2-day period (0.0, 0.2, 0.75, 1.0, 1.5 and 2.0 VSL) —48 challenges
- 6 different standard solutions; 2 μL injected for each challenge; blind, randomized challenges
- P&A study passed, UIFM of 14%
- PCAPP goal of LOQ < 0.2 VSL was met
- Accuracy improved significantly

Study Name: 20-23 Nov 2013 9361 9366 $n=0.76$



FAL: 0.8254

TAL: 0.7590

LOQ: 0.0137

DETL: 0.0070

DECL: 0.0136

UIFM: 13.69 %

UIAS: 12.05 %

Slope: 0.9493

Y-intercept: 0.0070

Percent Recovery: 95.63 %

Data Points: 48

$n = 0.76$



Investigation of Non-Linearity of the XSD



- OICO Applications Note
- OICO Technical Note
- Another detector based on the same technology as the XSD
- Statistical analysis of data for 125 challenge series of 17 different MINICAMS units at PCAPP
- Challenge data for HD and simulant chemicals at PCAPP and other sites
- Nine-month study for 3 MINICAMS units operating 24/7 at PCAPP





OICO Applications Note for the XSD



- Response linear and through origin (0,0) only at low masses
- Response non-linear at higher masses



XSD Reactor
Probe Assembly



OICO Applications Note 16561101, entitled "Using the Halogen Specific Detector (XSD™) as an Alternative to the ELCD in USEPA Methods"



OICO Notes on Operation of the XSD

September 2, 1994 – Dr. Rich Simon, Developer of the XSD at OICO




The detector response correlates with simple Langmuir adsorption (i.e. fixed number of adsorption sites, at equilibrium a fraction of the sites (Z) is occupied by adsorbed species, each site can adsorb only 1 atom, heat of adsorption is the 'same' for all sites (and independent of the fraction covered), and no interaction between species at different or neighboring sites). Let B equal the ratio of the adsorption rate to the desorption rate (i.e. commonly call the adsorption coefficient), then $Z = bP/(1+bP)$. At low partial pressures of atomic chlorine, $bP \ll 1$ so $Z = bP$ (i.e. adsorption is a linear function of the partial pressure of atomic chlorine). At higher pressures, $1 - Z = 1/(bP)$ -- the fraction of available sites becomes inversely proportional to the partial pressure of atomic chlorine.

Langmuir adsorption model

$$H = \frac{\alpha \text{ (VSL)}}{1 + \beta \text{ (VSL)}}$$

where α and β are constants

At PCAPP 4.05 ng Corresponds to 1.00 VSL*




*For a sample flow rate of 450 mL/min and sample period of 3 min




Another Detector (Not Used at PCAPP But) Based on the Same Technology as the XSD



- Manufactured by SRI Instruments (Las Vegas, Nevada)
- Sold as the Dry Electrolytic Conductivity Detector (DELCD)
- Response linear and through origin (0,0) only at low masses
- Response non-linear at higher masses




Data from Xiaojing Li (2009)




Data from Xiaojing Li (2009)

125 Data Sets Generated at PCAPP Were Analyzed Statistically



QQ Plot

Test of exponent values data set for normality



Distribution of exponent values determined

Statistical results:

Parameter	125 Data Sets ¹
Average exponent (n)	0.75
Avg correlation coefficient	0.996
Standard deviation	0.056
95% Confidence interval	0.64 to 0.86
Minimum exponent value	0.63
Maximum exponent value	0.90

¹Challenge series (125 each) for 17 different MINICAMS units, conducted Oct 9, 2013, through Jan 16, 2014.

Data Sets Reviewed from Various Sites

For HD and Other Chlorinated Chemicals and Another Detector Based on the Same Technology as the XSD



Test Site	Challenge Chemical	Equivalent HD Range, VSL ¹	Average Exponent	Instrument Type Tested	Number of Instruments	Number of Series ²
OICO (Texas)	Chloroform	0.99 to 4.94	0.68	GC with XSD	1	1
OICO (Texas)	Chlorobenzene	0.35 to 1.74	0.71	GC with XSD	1	1
OICO (Alabama)	Dichlorvos	0.09 to 0.67	0.92	XSD MINICAMS	8 ³	32
PCAPP	Dichlorvos	0.25 to 2.0	0.77	XSD MINICAMS	1	1
NRT Methodologies	Dichlorvos	0.5 to 2.1	0.83	XSD MINICAMS	2	2
PCAPP	HD	0.2 to 2.0	0.75 ± 0.11	XSD MINICAMS	17	125
BGCAPP	HD	0.5 to 2.0	0.78	XSD MINICAMS	4	16
ECBC	HD	0.5 to 2.0	0.88	XSD MINICAMS	4	16
ABCDF	HD	0.2 to 3.0	0.85	XSD MINICAMS	2	9
Univ of Waterloo ⁴	Aroclor 1254 (PCBs)	0.3 to 3.0	0.77	GC with DELCD	1	3
BEST Center	1-Chlorodecane	0.33 to 2.67	0.76	AMS with DELCD ⁵	10	52

¹HD-equivalent challenge range, based on chlorine content, a sample flow rate of 450 mL/min, and a sample period of 3 min.

²Number of challenge data series analyzed for each site.


³All PCAPP XSD MINICAMS units underwent P&A studies at OICO; only eight representative challenge data sets were reviewed.

⁴From Thesis for Master of Science in Chemistry, University of Waterloo, Xiaojing Li, 2009 (*Ref.5*).

⁵Automated air monitoring system (AMS) developed by Battelle (Aberdeen, Maryland) with a DELCD (same technology as the XSD).




Exponent Values Obtained During a 9-Month Study at PCAPP*



Serial No. 9358

PM included replacing the XSD reactor probe assembly

Serial No. 9616



*Three XSD MINICAMS units were operating 24/7 and challenged weekly in the range of 0.2 to 2.0 VSL.



Statistical Results for the 9-Month Study Compared to the 125 Data Sets Analyzed Previously



Parameter	125 Data Sets ¹	S/N 9358 ^{2,3}	S/N 9616 ^{2,3}	S/N 9000 ^{2,3}
Average exponent (n)	0.75	0.78	0.81	0.80
Avg correlation coefficient	0.996	0.999	0.999	0.999
Standard deviation	0.056	0.057	0.040	0.068
95% Confidence interval	0.64 to 0.86	0.67 to 0.89	0.73 to 0.89	0.66 to 0.94
Minimum exponent value	0.63	0.63	0.70	0.65
Maximum exponent value	0.90	0.92	0.89	0.90

¹Challenge series (125 each) for 17 different MINICAMS units, conducted Oct 9, 2013, through Jan 16, 2014.

²Challenges series for 3 different MINICAMS units conducted weekly from Feb 13 through Nov 20, 2014.

³Each of the MINICAMS units passed a P&A study after the end of the 9-month study.



Work Regarding the Response of the XSD MINICAMS Summarized in a Recent Report



**REVIEW OF THE TECHNICAL BASIS FOR AND THE IMPACT OF
CHANGING THE EXPONENT VALUE IN THE HD VSL METHOD FOR THE
XSD MINICAMS AT PCAPP**

Revision 0.0

Gary D. Sides, Ph.D.

Chemical Demilitarization Chief Scientist
Battelle Senior Research Leader

Claire E. Longo, M.S.

Battelle Statistician IV

February 3, 2015

Battelle Memorial Institute
Chemical Demilitarization Business Unit

REVIEWED BY

John W. Barton, Ph.D.

Chief Scientist, BGCAPP
Battelle Senior Research Leader

George M. Lucier, Ph.D.

Deputy Chief Scientist, BGCAPP

Carrie A. Kauffman, M.S.

Researcher, Battelle BEST Center



- The response of the XSD was previously assumed to be linear in demilitarization applications—in fact, the response of this detector correlates with Langmuir adsorption, which is non-linear.
- Literature from the manufacturer of the XSD (OICO) indicates that the detector is non-linear in the mass range of interest at PCAPP.
- Reviews of challenge data for eleven different combinations of sites and chemicals and two different detectors based on the same technology (XSD and DELCD), as well as additional literature reviews, confirmed the non-linearity of the XSD.
- The non-linear response of the XSD can be modeled over the concentration range of interest for the VSL method at PCAPP (0.2 to 2.0 VSL) using the equation $H = H_{CAL} (VSL)^n$ with the exponent value, n , set to 0.76.
- P&A studies conducted with the exponent, n , set to 0.76 yielded greatly improved results and limits of quantification less than the PCAPP target of < 0.2 VSL.
- Although some variation of the exponent values obtained were noted during a 9-month study using three MINICAMS units, these MINICAMS units passed P&A studies at the end of the test period with the exponent value at 0.76.